




Gear and spline inspection according to certified standards



Gear and spline technology with service



pure perfection

FRENCO

Deutscher Kalibrierdienst (DKD)
Akkreditierungsstelle

vertreten im

Deutschen AkkreditierungsRat



Akkreditierung

Die Akkreditierungsstelle des **Deutschen Kalibrierdienstes** akkreditiert hiermit

FRENCO GmbH

Verzahnungstechnik, Meßtechnik

Jacob-Baier-Straße 3

90518 Altdorf

nach DIN EN ISO/IEC 17025: 2005 für Kalibrierungen im Bereich / in den Bereichen:

dimensionelle Größen

Bestandteil der Urkunde ist: Anlage 04 (2 Seiten), 2007-06-26

DAR-Registriernummer: DKD-K-27401

Akkreditiert im DKD seit: 2000-04-17

Braunschweig, 2007-06-26

Dr.-Ing. Wolfgang Bosch
Leiter der Akkreditierungsstelle



Deutscher Kalibrierdienst
Akkreditierungsstelle

DKD

Anlage 04
vom 2007-06-26 zur Akkreditierungsurkunde des Kalibrierlaboratoriums

Registriernummer:
DKD-K-27401
Seite 1 von 2

bei
FRENCO GmbH Verzahnungstechnik, Meßtechnik
Jakob-Baier-Straße 3
90518 Altdorf

Telefon: 09187 9522-0
Telefax: 09187 9522-40
E-Mail: frengo@frengo.de

Leiter: Dipl.-Ing. (FH) Jan Kühl
Stellvertreter: Dipl.-Ing. (FH) Thomas Peter
Werner Schiller


Akkreditierung seit: 2000-04-17

Messgrößen:
Evolventennormale
Flankenliniennormale
Teilung und Rundlauf
Maß über Messkreis

Permanentes Laboratorium

General Information

For everything concerning gear and spline measuring FRESCO is the first address in Germany. With our equipment we are able to measure nearly all kinds of gear. A constant room temperature and top qualified employees guarantee highest precision. The traceability to PTB certified artefacts ensure a reliable appreciation of the measuring results. The FRESCO calibration laboratory is as a DKD – calibration station accredited. Now FRESCO is authorized to do a complete calibration of gears and spurs – and this as single calibration laboratory worldwide.



The Program

Wear measurement	Page 5
Inspection of parts	Page 6
Online certificate	Page 7
DKD-calibration	Page 8
UPM -certificate	Page 10
Deviation analysis	Page 11



Quality manager	Jan Kühl	09187/ 9522-34	jk@frenco.de
Second quality manager	Werner Schiller	09187/ 9522-15	ws@frenco.de
Distribution/ offers	Heinz Hasenkopf	09187/ 9522-11	hh@frenco.de
	Ursula Chwalka	09187/ 9522-30	uc@frenco.de

Wear measurement

Scope of services

All kinds of measuring instruments and measuring equipment for internal and external gears and splines like gages, master gears, one flank taper arbor, instruments for size inspection and others are tested. The standard wear measurement according to the directive VDL2618 sheet 1 includes the below listed services:

- **Checking the delivered number of pieces**
- **Cleaning**
- **Demagnetisation**
- **Optical check on damages**
- **Determination of the gear and spline data by dint of standards and an existing drawing or**
- **Creation of a drawing (additional charge)**
- **Remedy of small defects at the gear and spline**
- **Size inspection of the gear and spline by hand with Abbe-length measuring instrument or length measuring comparator**
- **Inspection of form like profile and helix trace pitch and runout on our measuring machines**
- **Issue of an inspection certificate**
- **Electronic archiving of the measuring results at FRENCO**
- **New marking (additional charge)**
- **Packing in a rust protective foil**

Note:

If the wear measurement has to be aborted due to a size being out of tolerance, costs will be reduced accordingly.

Delivery time

The delivery time is about 2 weeks after the arrival of the objects to be tested at FRENCO. The delivery time can be shortened in accordance with the above mentioned persons (additional charge). The delivery time for DKD-calibrations is 10 weeks after the arrival at FRENCO.

Documentation

For each measured object a inspection certificate is issued, which contains:

- the object data, the specified and the actual size and the final appreciation on the front page
- depending on the scope of the order the diagrammed measuring results of the inspection machines with profile and helix trace, pitch and runout on the inner pages
- an index of the used abbreviations, the traceability and the measuring uncertainty on the back page.

Delivery

We ask you to send the measuring instruments for FRENCO free of charge in a for measuring instruments reasonable packing. The delivery from FRENCO back to you is carried out by a parcel service (on request by one of your choice).

Online certificate


The screenshot displays the FRENCO website interface. At the top left is the logo "pure perfection" with five red stars above it, and "FRENCO" in large blue letters below. To the right, there is a banner for "The FRENCO IC-artefact - a new impulse!" featuring images of precision-ground gears and a shaft. Further right, it states "FRENCO GmbH - for more than 25 years an expert in the gear and spline measurement." Below the banner is a navigation menu with links: Home, Language (English), Search (GO!), Extended Search, Imprint, Terms & Conditions, About Frenco, News & Events, Products and Services, Download, DKD, and Contacts. A left sidebar lists: Brochures, Software demo versions, Inspection certificate (highlighted in red), LogIn Software, and LogIn Representatives. The main content area is titled "Inspection certificate" and shows a form for a "new condition" inspection. The form contains the following data:

name	sign	date	customer
Kindler		10.08.2005	
order number	record id	drawing number	
20068888	66099	46602	
customer part no.		serial part no.	
		03	
part name			
Hub simulator max ANSI B92.1 similar - z25 x DP24/48 x α45			

Below the form, there are input fields for "Order Number: 20068888" and "Record ID: 66099", followed by a "Find file" button and a "print" icon.

Every customer can retrieve the certificate for all wear inspections and inspections of parts calibrated by FRENCO from our homepage as pdf-file. A detailed description can be found on our homepage www.frenco.de.

Before destroying the paper certificate please save the record ID and the order number. Without these codes the online certificate cannot be retrieved.



DKD-calibration for artefacts...

The calibration laboratory of FRENCO is accredited by the DKD according to DIN/ISO 17025 for the calibration of all principal features (DKD-K-27401). Thus not only helical and involute artefacts can be calibrated but also pitch artefacts and gears with pitch, runout and size between measuring circles.

This accreditation is valid for measurement ranges, which are demanded most frequently. For details concerning our range of accreditation see


www.dkd.info/de/laboratorien.htm.

Artefacts not being within this range can be calibrated as working standard calibration.

As result you receive a calibration certificate containing explanations concerning the measuring process, specifications to determine the position of the object to be measured and –of course- the measuring uncertainty.





Traceability



...and IC-artefacts

The smallest ascertainable measuring uncertainty depends on the size of the artefact. For DKD-calibrations it ranges as follows:

	U (k=2)	U (k=2)	U (k=2)
Profile	F_{α} : 1.5 - 2.3 μm	$f_{h\alpha}$: 1.0 - 2.0 μm	$f_{f\alpha}$: 1.0 μm
Helix	F_{β} : 1.7 - 2.3 μm	$f_{h\beta}$: 1.3 - 2.1 μm	$f_{f\beta}$: 1.0 μm
Pitch/ runout	F_p : 1,1 μm	f_p : 0,9 μm	F_r : 1,1 μm
Size over measuring circles	DOB: 1,5 - 1,9 μm		



u_{PM} – Certificate for measuring instruments of series V

On request FRENCO issues an u_{PM} - certificate for all measuring instruments of the series V (instruments for size inspection). This certificate is laid out according to VDA 5 and contains the measuring uncertainty of the measuring instrument. That eases the determination of the measuring uncertainty of the inspection process according to VDA 5, because the value of the measuring uncertainty just has to be applied. Moreover characteristic values like T_{min} , C_{gk} and C_g , which are also an indicator for the quality of your measuring instrument, are listed on the certificate.



Uncertainty of the measuring instrument according to VDA 5

FRENCO GmbH Jakob-Baier-Strasse 3 90518 Altdorf Tel.: +49 9187 9522-0 Fax: -40 www.frenco.de

		UPM - Prüfcertifikat Bestimmung u_{PM} nach VDA 5 (1.2003) zur Bestimmung der Prüfprozesseignung		09.05.2006 Bearbeiter: Jürgen Stellweg																																																				
Zg.-Nummer: 48466 Teil-Nr.: 1 Typ: IVMF 1x1 Messspanne: Max= 25.8327 Min= 25.7877 T= 0.045		Auftraggeber: ABCDEF Auftragsnummer: 12345678																																																						
1. Schritt: Auflösung i.O.? Auflösung: 0.002 Toleranz Werkstück: 0.045 Auflösung in %: 4.44% i.O. Toleranzklasse: 11																																																								
2. Schritt: Einstellmeister u_{Kal} Messunsicherheit des Einstellmeisters lt. Prüfcertifikat: U= 2 μm Zg.-Nummer: 48466-11-03-00 k= 2 (u_{ref}) $u_{Kal} = 1.00 \mu m$																																																								
3. Schritt: Wiederholgenauigkeit u_w 25 Wiederholmessungen am Kalibrierung mit Induktivtaster in μm :																																																								
<table border="1"> <tr><td>1...5</td><td>0.0</td><td>-0.2</td><td>-0.1</td><td>-1.6</td><td>-0.1</td></tr> <tr><td>6...10</td><td>-0.9</td><td>-1.5</td><td>-0.4</td><td>-1.9</td><td>-0.6</td></tr> <tr><td>11...15</td><td>-0.6</td><td>-0.7</td><td>-0.9</td><td>-0.8</td><td>-0.4</td></tr> <tr><td>16...20</td><td>-0.8</td><td>-0.9</td><td>-0.6</td><td>-0.6</td><td>-0.7</td></tr> <tr><td>21...25</td><td>-0.8</td><td>-0.8</td><td>-0.7</td><td>-1.3</td><td>-1.0</td></tr> </table>					1...5	0.0	-0.2	-0.1	-1.6	-0.1	6...10	-0.9	-1.5	-0.4	-1.9	-0.6	11...15	-0.6	-0.7	-0.9	-0.8	-0.4	16...20	-0.8	-0.9	-0.6	-0.6	-0.7	21...25	-0.8	-0.8	-0.7	-1.3	-1.0																						
1...5	0.0	-0.2	-0.1	-1.6	-0.1																																																			
6...10	-0.9	-1.5	-0.4	-1.9	-0.6																																																			
11...15	-0.6	-0.7	-0.9	-0.8	-0.4																																																			
16...20	-0.8	-0.9	-0.6	-0.6	-0.7																																																			
21...25	-0.8	-0.8	-0.7	-1.3	-1.0																																																			
$s_w = 0.461049527$ Auflösung Ind.-Tast: 0.0001 Verhältnis: Auflösung / $s_w > 2$? NEIN dann $u_w = s_w$ (u_g) $u_w = 0.46 \mu m$ C_{gk} und C_g nach Bosch (Heft 10): $C_{gk} = 3.25$ $C_g = 2.71$																																																								
4. Schritt: Systematischer Fehler der Messvorrichtung u_{sys}																																																								
<table border="1"> <thead> <tr> <th>Nr.</th> <th>Sollwert in mm</th> <th>Istwert in mm</th> <th>Abweichung in μm</th> </tr> </thead> <tbody> <tr><td>1</td><td>25.8327</td><td>25.8328</td><td>0.1</td></tr> <tr><td>2</td><td>25.8327</td><td>25.8330</td><td>0.3</td></tr> <tr><td>3</td><td>25.8327</td><td>25.8328</td><td>0.1</td></tr> <tr><td>4</td><td>25.8327</td><td>25.8328</td><td>0.1</td></tr> <tr><td>5</td><td>25.8102</td><td>25.8100</td><td>-0.2</td></tr> <tr><td>6</td><td>25.8102</td><td>25.8100</td><td>-0.2</td></tr> <tr><td>7</td><td>25.8102</td><td>25.8101</td><td>-0.1</td></tr> <tr><td>8</td><td>25.8102</td><td>25.8100</td><td>-0.2</td></tr> <tr><td>9</td><td>25.7877</td><td>25.7875</td><td>-0.2</td></tr> <tr><td>10</td><td>25.7877</td><td>25.7876</td><td>-0.1</td></tr> <tr><td>11</td><td>25.7877</td><td>25.7876</td><td>-0.1</td></tr> <tr><td>12</td><td>25.7877</td><td>25.7876</td><td>-0.1</td></tr> </tbody> </table>					Nr.	Sollwert in mm	Istwert in mm	Abweichung in μm	1	25.8327	25.8328	0.1	2	25.8327	25.8330	0.3	3	25.8327	25.8328	0.1	4	25.8327	25.8328	0.1	5	25.8102	25.8100	-0.2	6	25.8102	25.8100	-0.2	7	25.8102	25.8101	-0.1	8	25.8102	25.8100	-0.2	9	25.7877	25.7875	-0.2	10	25.7877	25.7876	-0.1	11	25.7877	25.7876	-0.1	12	25.7877	25.7876	-0.1
Nr.	Sollwert in mm	Istwert in mm	Abweichung in μm																																																					
1	25.8327	25.8328	0.1																																																					
2	25.8327	25.8330	0.3																																																					
3	25.8327	25.8328	0.1																																																					
4	25.8327	25.8328	0.1																																																					
5	25.8102	25.8100	-0.2																																																					
6	25.8102	25.8100	-0.2																																																					
7	25.8102	25.8101	-0.1																																																					
8	25.8102	25.8100	-0.2																																																					
9	25.7877	25.7875	-0.2																																																					
10	25.7877	25.7876	-0.1																																																					
11	25.7877	25.7876	-0.1																																																					
12	25.7877	25.7876	-0.1																																																					
Max. Abweichung der Messvorr.: $F_{ges} = 0.3$ (ohne Feinzeiger) Faktor 0.58 (u_{BI}) $u_{sys} = 0.17 \mu m$																																																								
5. Schritt: Einfluss des Anzeigeelements u_{ind} Gesamtfehler des Feinzeigers lt. Kalibrierschein: 1.2 μm Dies entspricht: 1.2 μm Faktor 0.58 $u_{ind} = 0.70 \mu m$																																																								
$u_{PM} = \sqrt{u_{Kal}^2 + u_w^2 + u_{sys}^2 + u_{ind}^2}$ $u_{PM} = 1.39 \mu m$																																																								
$T_{min} = \frac{6 * u_{PM}}{G_p}$ $G_p = 0.2$ $T_{min} = 0.042$ $T > T_{min}$? JA																																																								




Deviation analysis

The measuring results are digitally filed and can be used in an analysis program for gears and splines. This program features:

- Changes of the position of axis
- Multiple measuring results can be overlaid for comparisons
- Display depending on the tooth thickness
- Correction of the settings of the manufacturing machine (profile grinding)
- Detection of disturbances in the production process,
- Simulation of changes of the datas of gears and splines
- Generation of new data of gear and spline

This analysis program helps to understand correlations, to appreciate disturbances and to detect failure causes. Moreover it features the possibility of a quick change of gear and spline data and many different ways of display them. The most important feature is, that the print of the analysis is like a measuring result and so the user can immediately understand it.



FRENCO Product Lines



Gear and spline high precision

Spline Gages
Master Gears, master wheels
Artefacts, masters
Profiled tools
Clamping systems
Gear and spline manufacturing




Instruments for size inspection Series V

Ball inserts and pins VRK
Instruments for rocking VA
Instruments with face stop VP
Indicating Gages VM
Variable 3-Disc Gages VD
Customized solutions VS



Rotation Measuring Systems

URM - K with balls and pins
URM - R with master wheels
EWP Single flank gear rolling
ZWP Double flank gear rolling
WS Gear Rollscan



Gear and spline inspection

DKD gear calibration
Gage wear inspection
Part inspections
Deviation analysis



Know-how transfer

Software for gear and spline calculating
Training, seminars and workshops
Consulting and calculations
Literature and documents
National and international standards



FRENCO

Frenco GmbH
gear + spline technology

Jakob-Baier-Straße 3

D 90518 Altdorf, Germany

Phone: +49 (0) 9187 9522 0

Fax: +49 (0) 9187 9522 40

E-Mail: frenco@frenco.de

Web: www.frenco.de